



# WATER TESTING SOLUTIONS

## APPLE CIDER VINEGAR ACIDITY TEST KIT

Tel: +44 (0)1903 962300  
enquiries@watertestingsolutions.com  
www.watertestingsolutions.com

Water Testing Solutions Ltd  
Unit C1, Rudford Ind Estate  
Ford Road  
Ford  
West Sussex  
BN18 0BD





# Apple Cider Vinegar Acidity Test Kit



In the UK, apple cider vinegar (ACV) must meet specific acidity standards, generally requiring a minimum of 5% acetic acid by volume. This acidity level is critical for preserving ACV's stability, antimicrobial qualities, and shelf life. The Food Standards Agency oversees these guidelines to ensure safe and high quality ACV production for consumers.

To support producers in meeting these standards, Water Testing Solutions offer a user-friendly drop test kit designed specifically for measuring the acidity of ACV 0-10%. This kit provides accurate results in just a few steps, allowing producers to verify that each batch meets regulatory requirements efficiently. Our drop test kit is ideal for quick, reliable testing on-site, ensuring compliance and quality control with minimal equipment and expertise needed.

Product Code	Equipment & Reagents
WTS-K102784	Apple Cider Vinegar Acidity Test Kit 0-10%
WTS-B102785	VA1 Acidity Indicator (30ml)
WTS-B102786	VA2 Acidity Titrant (65ml)
WTS-B100636	10ml Tube & Cap
WTS-B100619	Plastic 5ml Syringe

Tel: +44 (0)1903 962300

[enquiries@watertestingsolutions.com](mailto:enquiries@watertestingsolutions.com)

[www.watertestingsolutions.com](http://www.watertestingsolutions.com)